

THROUGH MAIL

OFFICE OF THE SUPERINTENDENT, V.S.S INSTITUTE OF MEDICAL SCIENCES &  
RESEARCH, BURLA, (SAMBALPUR)  
Mail ID: vimsarsuptd@gmail.com  
\*\*\*\*

To No. \_\_\_\_\_ /VIMSAR Dated, Burla, the \_\_\_\_\_ th October, 2022

The Deputy Director (Advt.)  
Information and PR Department, Orissa.  
Bhubaneswar.

Sub: Publication of Tender Call Notice.

Sir,

I am sending herewith the Tender Call Notice No. \_\_\_\_\_ /VIMSAR dt. \_\_\_\_\_ through mail with a request to publish it in at least two leading Odia daily News papers for its wide circulation and also necessary arrangement may kindly be made to supply a copy of the said news papers to the office of the Superintendent, VSS Institute of Medical Sciences & Research, Burla for record. The confirmation of the same may be sent through E.mail : vimsarsuptd@gmail.com.

The total expenditure will be come to Rs.15 lakhs (approximately)

Yours faithfully,

Encl: Tender Call Notice

  
Superintendent,  
VSS Institute of Medical Sciences &  
Research, Burla

Memo No. 13565 /VIMSAR Dated, Burla, the 06 /th October, 2022

Copy along with copy of Tender Call Notice forwarded to the Director, Printing and Stationary, Madhapatna, Cuttack-10 for favour of information.

Copy along with the copy of the Tender Call Notice forwarded to the Director of Medical Education and Training, Odisha, Bhubaneswar for favour of information.

Copy along with a soft copy forwarded to the District Informatics Officer, Collectorate, Sambalpur for information and necessary action with a request to float the tender call notice in the district website for its wide circulation.

Copy to the Dean & Principal, VSS Institute of Medical Sciences & Research, Burla/ Administrative Officer/Blood Bank Officer/Accounts Officer/Stores Medical Officer, VIMSAR, Burla for needful. They are requested to please remain present at the time of opening of the tenders on dt.11-11-2022 at 12-30 P.M.

  
Superintendent,  
VSS Institute of Medical Sciences &  
Research, Burla

E.mail: vimsarsuptd@gmail.com

GOVERNMENT OF ODISHA  
HEALTH & FAMILY WELFARE DEPARTMENT  
OFFICE OF THE SUPERINTENDENT, VSS INSTITUTE OF MEDICAL SCIENCES &  
RESEARCH, BURLA, (SAMBALPUR) ODISHA-768017

\*\*\*\*\*

No. 13563 /VIMSAR Dated, Burla, the 06<sup>th</sup> October, 2022

TENDER CALL NOTICE

Sealed Tenders are hereby invited by the undersigned from the reputed Manufacturer/Authorized Dealers/D.G.S. & D. Rate Contract holders and Registered firms under GST Act for supply of the following equipments required for the blood centre of this hospital, so as to reach this office of the Superintendent, VSS Institute of Medical Sciences & Research, Burla on or before dt. 10-11-2022 up to 4-00 P.M. through Regd.Post/ Speed Post/Couriers only and that will be opened in the next working day at 12.30 P.M.

1. Tube Sealer
2. Blood Storage Cabinet-600Ltr.

The detail terms and conditions of this Tender Call Notice along with the technical specifications of the above equipments can be downloaded from Dist. website [www.Sambalpur.nic.in](http://www.Sambalpur.nic.in)

The undersigned/committee reserves the full right to accept in full or reject any or all the tenders without assigning any reasons thereof.

*Clm*  
Superintendent,

VSS Inst. of Med.Sc.& Research, Burla  
*Clm*

**Product Name: Dielectric RF Tube Sealer**

Sl No	Specification
1	Should be a Heavy-duty radio frequency sealer
2	Should have Automatic detection of the tube by pressing of a lever which activates sensor
3	The equipment must not have warm up time before sealing
4	Should have separable rupture line to separate tube ends after sealing
5	Should have Switch mode power supply of uniform sealing irrespective of power supply variations
6	Standard blood bag tubes with tube diameter upto 6mm
7	Should have adjustable sealing time between 1-3 seconds.
8	RF signal should have applied only after the tube is fully squeezed
9	Should have removable splash guard to protect phlebotomist from all kinds of blood splashes
10	Should have Indications for Cover open, Ready, Seal and Power
11	The equipment must be protected against electric shock
12	Must be operational on 185W; standby 20W
13	Should have Easy to clean electrodes which are easily accessible & protected by cover
14	Should have CE, ISO 9001 :2008, ISO13485, EURoHS
15	Should have a SMPS to withstand input voltage fluctuations
16	Should be easy to handle and must weight approximately 5.25 Kg

*Am*  
5-10-22

HOD Transfusion Medicine  
VSS Institute of Medical Science  
and Research, Burla, SOP.

**Blood Bank Refrigerator 600 Litres** ✓

Sl no	Specification
1	Should have a chamber temperature range of 2°C to 6°C
2	Should comply IEC safety /EMI/EMC standards, ISO 9001 ,ISO 13485 certified and CE marked with Class II :
3	Should have a total storage capacity of 510 nos of 350ml blood bags
4	Should have a corrosion free Interior of SS 304 gradewhich helps in easy cleaning
5	Should have a 18 swg CRCA/HDGI Sheet - powder coated
6	Should have built in microcontroller based temperature recorder and controller unit (TRCU) positioned a eye level for better visibility and monitoring
7	Must have six inch, 7-day, ink-less pressure sensitive circular chart recorder and accuracy should be +/- degree celsius
8	Should have triple pane glass door which insulates the refrigerator from atmospheric temperature and bette visibility of blood bags with key lock.
9	Must have an internal evaporator fan for uniform temperature maintenance inside the chamber. It shoul switch off automatically whenever the door is opened
10	Should have perforated sliding Stainless steel tray which allows bags to be placed upright with sufficien airspace to reduce " Sardine effect"
11	Should have 5 stainless steel slide out trays
12	Should have separate inner acrylic door with magnetic latch ensures minimal loss of cooling
13	Should have hotline around the mouth of cabinet to prevent moisture condensation
14	Should have internal stabilizer which ensures steady supply voltage for the whole blood storage cabinet an prevents voltage fluctuations
15	Should have a factory calibrated encapsulated digital sensor dipped in liquid medium to measur temperature at an accuracy of ±0.5°C
16	Should have flicker free CFL lamp for uniform lighting and better visibility of samples inside the cabinet
17	Should include caster wheels as a standard feature
18	Should have refrigeration system "On" indicator provided as a standard feature
19	Should have audible and visual high and low temperature alarms as a standard feature
20	Should have an alarm silence button
21	Should have a power fail alarm as a standard feature
22	Must incorporate a inbuilt battery back up of 2 hours to ensure continuous operation of the TRCU durin power failure
23	Should have display of 4*7 Segment LED with 0.5°C display resolution
24	Must utilize non-CFC, commercially available refrigerant
25	Must keep the refrigerator free of frost without elevating the chamber temperature
26	Must have a cabinet with a clear powder coated finish to guard against rust and corrosion
27	There must be eye level indications of power/Line in from inbuilt stabilizer
28	Chart change and Chart set options in TRCU must be available
29	Must have MCB protection in the input power line.

Q2/13+14

DECLARATION FORM

I/We \_\_\_\_\_  
having My/our \_\_\_\_\_ office at declare  
that I/We have carefully read all the conditions of the Superintendent, VSS Institute  
of Medical Sciences & Research, Burla for the supply of Consumables & others  
for the tender period of one year. I will abide by all conditions set forth in the tender  
paper for the year, 2022-23.

I/We further declare that I/We possess valid manufacturing  
License/Distributor Licence Bearing No. \_\_\_\_\_ Valid Up to \_\_\_\_\_

I/We \_\_\_\_\_ do  
hereby declared that I/we will supply the items as per the terms & conditions of the  
tender call notice. Further the rates quoted by me for this tender call notice are not  
more than the prevailing market price and if detected later on the differential  
amount will be recovered from our company.

I/We promise to give \_\_\_\_\_ years warranty/guarantee on the items (Items  
to be mentioned) supplied by me against any manufacturing defect/rusting/break  
down. I will replace the item without any extra cost from the purchaser failing  
which the cost of the items will be recovered from my security/performance deposit  
and EMD as the case may be. We will charge no advance payment towards the  
replacement of any part/parts during the Warranty/Guarantee Period.

Signature of the bidder:

Date :

Name & Address of the Firm:

Affidavit before Executive Magistrate/Notary Public.