

8

E.mail: vimsarsuptd@gmail.com

**GOVERNMENT OF ODISHA
HEALTH & FAMILY WELFARE DEPARTMENT
OFFICE OF THE SUPERINTENDENT, VSS INSTITUTE OF MEDICAL SCIENCES &
RESEARCH, BURLA, (SAMBALPUR) ODISHA-768017**

No. 9834 /VIMSAR Dated, Burla, the 06 th April, 2023

TENDER CALL NOTICE

Sealed Tenders are hereby invited by the undersigned from the reputed Manufacturer/Authorized Dealers & Registered firms under GST Act for supply of Manual Air Oxygen Blender required for SNCU pf Pediatric department of this hospital, so as to reach this office of the Superintendent, VSS Institute of Medical Sciences & Research, Burla on or before dt. 03.05.2023 up to 1-00 P.M. through Regd.Post/ Speed Post/Couriers only and that will be opened in the next working day at 11.30 A.M.

The detail technical specification along with terms and conditions of this Tender Call Notice can be downloaded from Dist. website www.Sambalpur.nic.in

The undersigned/committee reserves the full right to accept in full or reject any or all the tenders without assigning any reasons thereof.


o/c
Superintendent,
VSS Inst. of Med.Sc.& Research, Burla

[Signature]

TERMS AND CONDITIONS

1. The tenderer should be a Registered Firm under GST Act.
2. The rate should be inclusive of all taxes except GST. The GST charges should be clearly mentioned.
3. The delivery of the item should be made in the main medical store of this Hospital in good condition which will be installed by the firm in the SNCU of this hospital.
4. Any conditional Tender will not be accepted.
5. No advance payment will be made against execution of order. Payment will be released only after successful supply & satisfactory installation and certification.
6. The rate should be valid for one year from the date of approval..
7. Original documents may be produced for verification if asked for.
8. If any information/documents furnished by the tenderers with the tender papers are found to be misleading or in correct, at any stage their tenders will be cancelled.
9. The make & model of the quoted item must be mentioned in the tender. The warranty/guarantee of the item should be mentioned clearly.
10. The product must be of good quality and also ISI/CE/BIS/ISO certified. The documents thereof should also be produced along with the tender.
11. The tenders should be furnished in time with sealed cover duly super scribed as **"TENDER FOR SUPPLY OF MANUAL AIR OXYGEN BLENDER AND DUE DATE IS DT. 03.05.2023 UP TO 01.00 P.M.**
12. The tenders will be received up to 1.00 P.M. of dt.03.05.2023 through Regd.Post/Speed Post/Courier only to the office of the Superintendent, VSS Institute of Medical Sciences & Research, Burla. The tenders received beyond the scheduled date & time will not be entertained. The postal/courier delay is the responsibility of the tenderers. The tenders will be opened on dt.04.05-2023 by a Committee at 11.30 A.M. The tenderer may attend while opening the tenders.
13. The cost of the Tender call Notice is Rs.2000/- which is non refundable. The cost of tender should be in shape of Bank Draft drawn in favour of the Superintendent, VIMSAR, Burla to be enclosed with the tender documents.
14. The following documents will be enclosed in the tender paper.
 - (i) Attested copy of GST Registration Certificate.
 - (ii) Attested copy of the PAN Card.
 - (iii) Rs.8000/-in shape of NSC/Bank Draft/FDR duly pledged in favour of the Superintendent, VSS Institute of Medical Sciences & Research, Burla should be deposited as EMD which will be refunded after selection of the rate. If any exemption to be availed than the authentic supporting documents to be enclosed otherwise the tender may be cancelled.
 - (iv) The authorization letter of the Principal Firm should be enclosed in the Tender with due attested, if the firm is an authorized dealer.
 - (v) The Bank Draft of Rs.2000/-drawn in favour of the Superintendent, VIMSAR, Burla towards cost of tender paper to be enclosed in the tender. If any exemption to be availed than the authentic supporting documents to be enclosed otherwise the tender may be cancelled.
 - (vi) An undertaking regarding "The quoted rates are not more than the prevailing market price" and will be valid for one year.
 - (vii) Catalogue of the item.
 - (viii) The required documents like ISO/BIS/CE/ISI certificates of the concerned manufacturers should be enclosed in the tender for selection of the quality items.

The undersigned/committee reserves the full right to accept in full or part or reject any or all the tenders without assigning any reasons thereof. The selection will be made on quality basis wherever necessary.



 Superintendent,
 VSS Inst.of Med.Sc.& Research ,Burla

TECHNICAL SPECIFICATION OF ,MANUAL AIR OXYGEN BLENDER

FiO2	21%-100%
Flow Control	1-10 LPM on both ports
Pressure	Air/Oxygen @0.3-0.4Mpa
Alarm	When gas supply pressure difference > 0.1MPa
Alarm Noise	> 57dB(A)
Number of Output	Dual
Accuracy of FiO2	±3%
Bleed Flow	3LPM
<ul style="list-style-type: none">➤ Company shall provide the wheel stand with wheel lock facility & 3 mounts with each blender➤ The air & oxygen hose pipe should be at least 2 meter in length with different colors.➤ There should be provision for the air hose pipe to be connected to medical air compressor.	

o/c




Superintendent
VSS. Institute of Medical Sciences & Research
Burla, Sambalpur