

E.mail: vimsarsuptd24@gmail.com

GOVERNMENT OF ODISHA
HEALTH & FAMILY WELFARE DEPARTMENT
OFFICE OF THE SUPERINTENDENT, VSS INSTITUTE OF MEDICAL SCIENCES &
RESEARCH, BURLA, (SAMBALPUR) ODISHA-768017


No. 10789 /VIMSAR ***** Dated, Burla, the 05th ^{Aug} July, 2024

TENDER CALL NOTICE

Sealed Tenders are hereby invited by the undersigned from the reputed Manufacturer/Authorized Dealers and Registered firms under GST Act for supply of Blood Collection Monitor required for the Blood Centre of the VIMSAR, Burla, so as to reach this office of the Superintendent, VSS Institute of Medical Sciences & Research, Burla on or before dt.30-08-2024 up to 5.00 P.M. through Regd.Post/ Speed Post/ Couriers only and that will be opened on 31-08-2024 at 11.30 A.M. The delay in receipt of tender paper is the responsibility of the tenderer.

The detail technical specifications of the above equipment along with terms and conditions of this Tender Call Notice can be downloaded from Dist. website www.Sambalpur.nic.in


The undersigned/committee reserves the full right to accept in full or reject any or all the tenders without assigning any reasons thereof.


Superintendent,
VSS Inst. of Med.Sc.& Research, Burla


TERMS AND CONDITIONS

1. The tenderer should be a Registered Firm under GST Act.
2. The rates should be inclusive of all taxes except GST. The GST charges should be clearly mentioned.
3. The delivery of the equipment should be made in the Blood Centre of this Hospital in good condition & should be installed in the designated places.
4. Any conditional Tender will not be accepted.
5. No advance payment will be made against execution of order. Payment will be released only after successful supply & satisfactory installation and certification.
6. The rates should be valid for one year from the date of approval..
7. The exemption of Tender Cost or EMD will be made as per the rule of Govt. Supporting documents will be enclosed in the tender for verification.
8. Original documents may be produced for verification if asked for.
9. If any information/documents furnished by the tenderers with the tender papers are found to be misleading or in correct, at any stage their tenders will be cancelled.
10. The make & model of the quoted equipment must be mentioned in the tender. The technical aspects of the products should be similar to the technical specifications mentioned in the tender call notice. The warranty/guarantee of the equipment along with after sales & service to be mentioned clearly.
11. The product must be of good quality and also ISI/CE/BIS/ISO certified. The documents thereof should also be produced along with the tender.
12. The tenders should be furnished in time with sealed cover duly super scribed as
"TENDER FOR SUPPLY OF BLOOD COLLECTION MONITOR FOR BLOOD CENTRE AND DUE DATE IS DT. 30-08-2024 UP TO 5.00 P.M.
13. The tenders will be received up to 5.00 P.M. of dt.30-08-2024 through Regd.Post/Speed Post/Courier only to the office of the Superintendent, VSS Institute of Medical Sciences & Research, Burla. The tenders received beyond the scheduled date & time will not be entertained. The postal/courier delay is the responsibility of the tenderers. The tenders will be opened on dt. 31-08-2024 by a Committee at 11.30 A.M. The tenderer may attend while opening the tenders but that will not be a bar.
14. The cost of the Tender Call Notice is Rs. 2000/- which is non refundable. The cost of tender should be deposited in shape of Bank Draft drawn in favour of the Superintendent, VIMSAR, Burla.
15. The following documents will be enclosed in the tender paper.
 - (i) Copy of GST Registration Certificate.
 - (ii) Copy of the PAN Card.
 - (v) Rs.10,000/-in shape of NSC/Bank Draft/FDR duly pledged in favour of the Superintendent, VIMSAR, Burla should be deposited as EMD which will be refunded after selection.
 - (vi) The authorization letter of the Principal Firm should be enclosed in the Tender if the firm is an authorized dealer.
 - (x) The tender paper cost of Rs.2000/- will be enclosed as Bank Draft duly pledged in favour of the undersigned.
 - (xi) The documents for exemption of EMD or Tender cost should be enclosed if availed.
 - (xii) An undertaking regarding "The quoted rate is not more than the prevailing market price" and will be valid for one year.
 - (xiii) Catalogue of the equipment.
 - (xiv) The required documents like ISO/BIS/CE/ISI certificates of the concerned manufacturers should be enclosed in the tender for selection of the quality products.
 - (xv) The user list of the said equipments should also be enclosed with the tender.

The undersigned/committee reserves the full right to accept in full or part or reject any or all the tenders without assigning any reasons thereof. The selection will be made on quality basis wherever necessary.


Superintendent,
VSS Inst.of Med.Sc.& Research ,Burla

Product Name: Blood Collection Monitor


Sl no	Specification
1	Should comply IEC safety / ISO 9001:2015, ISO 13485:2016 certified and CE marked as Class Im (Measuring)
2	Must have a weighing range of 50-500 ml
3	Must have Automatic tare to zero for the bag weight
4	The default volume must be adjustable
5	Should clamp Automatically at termination of preset volume collection
6	Should have option for manual clamp for usage in case of emergency
7	Must have a weighing accuracy +/-2%
8	Must have easily detachable tray for cleaning and disinfection which are compatible with all bag systems
9	Must have 20 x 4-line character Backlit LCD display of programmed volume, collected volume, flow rate, main battery,
10	Alarm /Indications for high/low flow indication, end of collection, power failure and battery low
11	Must have indication of time taken for collection
12	Must have indication with alarm if blood flow rate is high or low
13	Must have continuous display of collected volume, flow and time during collection
14	Trays Should oscillate continuously at 12+/- 2 rpm during collection
15	Should operate on mains as well as rechargeable battery. Lithium Ion Battery with 12 hours battery backup/ 90 to 100 collections assuming an average 8-minute donation time
16	Pause facility to pause during collection
17	350/450 ml easy selection switch
18	Should have following alarm & indications: Audible/visible: Flow rate below 20 mL/min or above 180 mL/min Audible/visible: Power failure Audible/visible: Battery low Audible/visible: Automatic clamping when flow rate is sustained at less than 20 mL/min for more than 2 minutes Audible/visible: End of collection
19	Should have external dimension 190 mm x 500 mm x 185 mm (+/- 5mm Tolerance)
20	working environment: Temperature: 10 °C to 40 °C Relative humidity: 30% to 75%
21	Should have wight check mode.


Date: 5/07/24
 Medical Officer I/C
 Odisha Blood Centre
 VIMSAR, BURLA

Product Name: Blood Collection Monitor

Sl no	Specification
1	Should comply IEC safety / ISO 9001:2015, ISO 13485:2016 certified and CE marked as Class Im (Measuring)
2	Must have a weighing range of 50-500 ml
3	Must have Automatic tare to zero for the bag weight
4	The default volume must be adjustable
5	Should clamp Automatically at termination of preset volume collection
6	Should have option for manual clamp for usage in case of emergency
7	Must have a weighing accuracy +/-2%
8	Must have easily detachable tray for cleaning and disinfection which are compatible with all bag systems
9	Must have 20 x 4-line character Backlit LCD display of programmed volume, collected volume, flow rate, main battery,
10	Alarm /Indications for high/low flow indication, end of collection, power failure and battery low
11	Must have indication of time taken for collection
12	Must have indication with alarm if blood flow rate is high or low
13	Must have continuous display of collected volume, flow and time during collection
14	Trays Should oscillate continuously at 12+/- 2 rpm during collection
15	Should operate on mains as well as rechargeable battery. Lithium Ion Battery with 12 hours battery backup/ 90 to 100 collections assuming an average 8-minute donation time
16	Pause facility to pause during collection
17	350/450 ml easy selection switch
18	Should have following alarm & indications: Audible/visible: Flow rate below 20 mL/min or above 180 mL/min Audible/visible: Power failure Audible/visible: Battery low Audible/visible: Automatic clamping when flow rate is sustained at less than 20 mL/min for more than 2 minutes Audible/visible: End of collection
19	Should have external dimension 190 mm x 500 mm x 185 mm (+/- 5mm Tolerance)
20	working environment: Temperature: 10 °C to 40 °C Relative humidity: 30% to 75%
21	Should have wight check mode.

ok


Superintendent
VSS. Institute of Medical Sciences & Research
Burla, Sambalpur-768017


Medical Officer I/C
Odisha Blood Centre
VIMSAR, BURLA