

Through mail

OFFICE OF THE SUPERINTENDENT, V.S.S INSTITUTE OF MEDICAL SCIENCES &
RESEARCH, BURLA, (SAMBALPUR)
E.mail – vimsarsuptd24@gmail.com

To No. _____ /VIMSAR Dated, Burla, the _____ th May, 2025

The Deputy Director (Advt.)
Information and PR Department, Orissa.
Bhubaneswar.

Sub: Publication of Tender Call Notice.

Sir,

I am sending herewith the Tender Call Notice No. 6503 /VIMSAR dt. 06-05-2025 through mail with a request to publish it in at least two leading Odia daily News papers & one English daily for its wide circulation and also necessary arrangement may kindly be made to supply a copy of the said news papers to the office of the Superintendent, VSS Institute of Medical Sciences & Research, Burla for record. The confirmation of the same may be sent through E.mail : vimsarsuptd24@gmail.com.

The total expenditure will be around Rs.20 lakhs (approximately)

Yours faithfully,

Encl: Tender Call Notice


Superintendent,
VSS Institute of Medical Sciences &
Research, Burla

Memo No. 6505 /VIMSAR Dated, Burla, the 6th May, 2025

Copy along with copy of Tender Call Notice forwarded to the Director, Printing and Stationary, Madhupatna, Cuttack-10 for favour of information.

Copy along with the copy of the Tender Call Notice forwarded to the Director, VIMSAR, Burla for favour of information.

Copy along with a soft copy forwarded to the District e-Governance Manager (DeGM), Sambalpur for information and necessary action with a request to float the tender call notice in the district website for its wide circulation.

Copy to the Dean & Principal, VSS Institute of Medical Sciences & Research, Burla/ Registrar (Administration)/HOD Transfusion Medicine /Accounts Officer/Stores Medical Officer/Medical Officer i/c Blood Centre, VIMSAR, Burla for needful. They are requested to please remain present at the time of opening of the tenders on dt.06-06-2025 at 11-30 A.M.


Superintendent,
VSS Institute of Medical Sciences &
Research, Burla

E.mail: vimsarsuptd24@gmail.com

HEALTH & FAMILY WELFARE DEPARTMENT, GOVT. OF ODISHA
OFFICE OF THE SUPERINTENDENT
VSS INSTITUTE OF MEDICAL SCIENCES & RESEARCH, BURLA (SAMBALPUR)
ODISHA-768017
Mail id: vimsarsuptd24@gmail.com

No. 6503 /VIMSAR Dated, Burla, the 6 th May, 2025

TENDER CALL NOTICE

Sealed Tenders are hereby invited in two bid systems by the undersigned from reputed Manufacturers/Authorized Dealers/D.G.S.&D. Rate Contract holders/ EPM Rate contract holders and Registered firms under GST Act with validity for one year for supply of 1. Blood Storage Cabinet 2. Hot Air Oven to the Blood Centre of this hospital so as to reach the office of the Superintendent, VSS Institute of Medical Science & Research, Burla on or before dt.05-06-2025 up to 1.00 P.M. through Regd.Post/Speed Post only and this will be opened at 11.00 A.M. in the next working day. The tenderer may attend the office at the time of opening of the tenders. The delay in receipt of tender paper is the responsibility of the tenderers.

The detail terms and conditions along with technical specifications of the equipments can be downloaded from Dist. website sambalpur.odisha.gov.in & Institute website www.vimsar.ac.in

The undersigned/committee reserves the full right to accept in full or reject any or all the tenders without assigning any reasons thereof.


Superintendent,
 VSS Institute of Medical Sc.& Research, Burla

TERMS AND CONDITIONS OF THE TENDER CALL NOTICE

1. Covers containing the tenders & documents should be superscribed as "Tender for supply of Equipments for Blood Centre" duly sealed, containing both the covers of Technical bid & Price bid inside it.
2. The cost of Tender paper is Rs.2000/-. The tenderers who wants to purchase tender documents from the office of the Superintendent, VSSIMS&R, Burla will deposit Rs.2000/- (Rupees Two thousand) only in the User's fee collection counter of this hospital during office hours on proper application through their authorized representative and enclose the red copy of the money receipt in their tender documents or can be downloaded from Dist. website Sambalpur.odisha.gov.in and Institute website www.vimsar.ac.in and tender can be submitted with a demand draft of Rs.2000/-(Rupees Two thousand) only drawn in favour of Superintendent, VSS Institute of Medical Sciences & Research, Burla towards the cost of tender paper with their tender. The cost of tender paper is not refundable.
3. The rate of the equipments should be quoted in Indian Rupees in the Tender and be submitted in the Proforma of price bid only. The validity of the offer should be one year.
4. The installation of the Equipments should be made in the Blood Centre of this Institution at their own cost and risk of the firm. The demonstration along with training if needed should be given on free of cost. The Instruments & equipments having best specifications & performance may be chosen. The documents relating to the quality of the equipments are to be given with the tender.
5. Any condition for supply & installation of Inst.& Equipments through Bank or Payment through letter of credit or advance payment will not be entertained. In-complete tenders as per the price bid/Technical bid or terms & conditions may lead for cancellation.
6. The approved firm is to furnish Security Deposit @ 10% of the face value of the purchase order duly pledged in favour of the Superintendent, VSS Institute of Medical Sciences & Research, Burla before supply which will be returned after successful Warranty/ Guarantee period. One Bi-Party agreement is to be made before execution of the order. The warranty /Guarantee period of the Inst.& Equipments should not be less than 2(two) years.
7. The Tenderer should mention the Warranty/Guarantee period of the equipments and an undertaking to maintain the equipments for at least 7(Seven) years including Warranty/Guarantee period and should enter into Annual Maintenance Contract (AMC or CMC) for which an agreement with some extra terms and conditions with non-judicial Stamp/Court Paper will be made if needed. No advance payment will be made towards repair or supply of spare parts for maintenance of the Equipments.
8. The tenderer must furnish the details of infrastructures required for installation of the equipments.
9. The tender submitted by the firms should be properly page marked.
10. The Manufacturing firms directly participating in this Tender Call Notice can not authorize any agent for their same products against this Tender Call Notice.
11. If the equipments show frequent troubles during warranty/Guarantee period then the defective spares will be replaced with new one at their own cost and risk within one month from the date of receipt of intimation letter.
12. The tender must be sealed and should be superscribed as "Tender for Supply of equipments to the Blood Centre, VIMSAR,Burla & Due Date is dt.05-06-2025 containing both the sealed covers of Technical bid & Price bid.

13. The last date of receipt of tender is **05-06-2025** up to 1.00 P.M. Beyond this the tenders will not be accepted. The tenders should be reached through Speed Post/ Regd.Post only. The tenders will be opened at 11.00 A.M, in the next working day by a Committee. The tenderer or their authorized representatives may attend this office while opening. But this is not a bar to the Committee to open the tenders or carry out subsequent tender process if the tenderer or authorized representatives cannot attend this office while opening.

14. **Cover & documents in the Technical Bid:** The following documents should be furnished in a sealed cover with an Index page (**Annexure-II**) failure to furnish any document, their tender will not be considered for evaluation.

(i) Photocopies of GST Registration Certificate, Income Tax Clearance Certificate (ITCC), GST Returns & PAN card.

(ii) Earnest Money Deposit (EMD) of Rs.20,000/- in shape of Fixed Deposit Receipt(FDR)/ NSC/Bank Draft duly pledged in favour of the Superintendent, VSS Institute of Medical Sciences & Research, Burla should be enclosed with the tender which will be returned after supply of the & Equipments by the selected firms. In case of others, this will be returned after selection of the rates & firms.

(iii) The Original Money Receipt of Rs.2000/- obtained towards cost of tender paper duly deposited in the users counter of the hospital should be enclosed with the tender or bank draft of Rs.2000/- drawn in favour of Superintendent, VIMSAR, Burla will be enclosed.

(iv) The firm should enclose the valid authorized dealership certificate along with the tender for the items quoted. No firm will be selected if the firm will not be an authorized dealer.

(v) The tenderer shall furnish the detail Literature/Brochures/Operating Manuals along with Technical details and Catalogues of the quoted items along with their tender.

(vi) The declaration form in **Annexure-I** must be signed by the tenderer before Notary Public/Executive Magistrate & must be enclosed with the tender in original.

(vii) The copy of the relevant up to date ISI/BSI/ISO certificate CE/DVN/GMP of the equipments and firms should be furnished along with the tender. If any other documents regarding quality of the above equipments should also be enclosed for evaluation.

(viii) The user list along with the copy of the orders if received towards the quoted items only for last two years along with their performance certificate of last installations should be enclosed with the tender.

15. **Documents be submitted in the cover of Price Bid:** The tenderers should submit the price bid duly filled in and signed and sealed as per **Annexure-III** giving the rates for the items, which should be inclusive of all taxes & incidental charges & excluding GST. The extra taxes should be clearly mentioned in the price bid. The cost of comprehensive/Annual maintenance charges (CMC/AMC) beyond the warranty/ Guarantee period and the cost of the Turn Key job with details, which it is applicable are to be indicated.

The tenderers are required to furnish the Technical Bid and Price Bid in **Annexure-II** and **Annexure-III** respectively in two separate sealed covers, which should be superscribed as cover containing price Bid and cover containing Technical Bid respectively. The cover containing the Technical Bid will be opened first for evaluation. The cover containing the price bid of the tenders who will be qualified after evaluation of the technical bids will be opened thereafter.

16. The undersigned/Committee reserves the full right to accept in full or part or reject any or all the tenders without assigning any reasons thereof. The undersigned/Committee is not bound to accept the lowest quoted rate of tenders considering the technical aspects and actual need of the blood centre for better patients care.


Superintendent,
YSS Institute of Medical Sc. & Research, Burla

ANNEXURE-I

DECLARATION FORM

I/We _____ having
My/our _____ office at declare that I/We
have carefully read all the conditions of the Superintendent, VIMSAR, Burla for the
supply & Installation of Equipments for the tender period of one year. I will abide by all
conditions set forth in the tender paper for the year, 2025-26.

I/We further declare that I/We possess valid manufacturing Licence/ Distributor
Licence Bearing No. _____ Valid Up to _____

I/We _____ do hereby
declared that I/we will supply & Install the Equipments as per the terms & conditions of
the tender call notice. Further the rates quoted by me for this tender call notice are not
more than the prevailing market price and if detected later on the differential amount will
be recovered from our company.

I/We promise to give _____ years warranty/guarantee on the items supplied by
me against any manufacturing defect/rusting/break down. I will replace the item
without any extra cost from the purchaser failing which the cost of the
instrument/Equipment will be recovered from my security/performance deposit and EMD
as the case may be. We will charge no advance payment towards the replacement of
any part/parts during the Warranty/Guarantee or during the AMC/CMC period for
functioning of the Equipments.

Signature of the bidder:

Date :

Name & Address of the Firm:

Affidavit before Executive Magistrate/Notary Public.

ANNEXURE-II

SCHEDULE OF REQUIREMENTS WITH TECHNICAL SPECIFICATIONS FOR
HOT AIR OVEN & BLOOD STORAGE REFRIGERATOR REQUIRED FOR THE BLOOD
CENTRE, VIMSAR, BURLA

Sl.No.	Item No. as per Tender document	Name of the item	Strength/Size/specification make and Model	Approximate tender quantity per year in absolute Number	Sample required
1.	2.	3.	4.	5.	6.

ANNEXURE-III

MODEL TENDER FORMAT (PRICE SCHEDULE)

Sl. No.	Item No.	Name of the item	Strength/Size & specification make & Model	Price per each unit which includes excise duty, packing forwarding warranty & maintenance, insurance, transportation & excludes sales tax and entry tax & customs duty (Imported items for each item.	AMC/CMC for the life span period	Tender Quantity	Total value
1.	2.	3.	4.	5.	6.	7.	8

Place:

Date:

Signature of the Bidder;

Name:

Address:

TECHNICAL SPECIFICATIONS OF HOT AIR OVEN

1. Capacity Ltr 200 and above (100Kg)
2. Inner Size (WxDxH) cm specify
3. Heater Load Kg/ Hr specify
4. No. Of Shelves nos 6-10
5. Temperature Range deg C up to 300
6. Temperature Uniformity : +/-2 deg C
7. Insulation- specify 9 safety Device- Thermostat/ specify
8. Type- Forced Convection System/ Others
9. Temperature Controller – Digital Temperature Controller cum indicator.
10. Inner & Outer Body- Made out of Stainless steel
11. Electrical Supply- Specify
12. Essential Spares- Specify
13. Total Connected Load Kw specify
14. Training: One day Training should be provided after installation.

TECHNICAL SPECIFICATIONS OF BLOOD STORAGE REFRIGERATOR

1. Should have a chamber temperature range of 2° C to 6° C.
2. Should comply IEC safety/ EMI/ EMC standards, ISO 9001, ISO 13485 certified and CE marked with class II a.
3. Should have an effective storage volume of 600 L (+ 20 liters) with a total storage capacity of 300 numbers of 450ml blood bags.
4. Should have built in microcontroller based temperature recorder and controller unit (TRCU) positioned at eye level for better visibility and monitoring.
5. Should have six inch, 7day, ink less pressure sensitive circular chart recorder and accuracy should be +/-1 degree Celsius.
6. Should have three pane glass doors which insulate the refrigerator from atmospheric temperature and better visibility of blood bags with key lock.
7. Should have an internal evaporator fan for uniform temperature maintenance inside the chamber. It should switch off automatically whenever the door is opened.
8. Should have mesh type sliding stainless steel tray which allows bags to be placed upright with sufficient air space to reduce "Sardine effect".
9. Should have 5 stainless steel slide out trays with a capacity of 60 to 70 nos of 450ml blood bags per tray.
10. Should have separate/ fixed inner acrylic door with magnetic latch ensures minimal loss of cooling.
11. Should have hotline around the mouth of cabinet to prevent moisture condensation.
12. Should have internal stabilizer which ensures steady supply voltage for the whole blood storage cabinet and prevents voltage fluctuations.
13. Should have a factory calibrated encapsulated digital sensor dipped in liquid medium to measure temperature at an accuracy of + 0.5°C.
14. Should have flicker free LED lamp for uniform lighting and better visibility of samples inside the cabinet.
15. Should include caster wheels as a standard features.
16. Should have refrigerator system "On" indicator provided as standard features.
17. Should have audible and visual high and low temperatures alarms as a standard features.
18. Should have an alarm silence button.
19. Should have power fail alarm as standard feature.
20. Should incorporate a inbuilt battery backup of 2 hours to ensure continuous operation of the TRCU (Temperature Recording Control Unit) during power failure.
21. Should have display of 4*7 segments LED with 0.5°C display resolution.
22. Should utilize non CFC, commercially available refrigerant.
23. Should keep the refrigerator free of frost without elevating the chamber temperature.
24. Should have a cabinet with a clear powder coated finish to guard against rust and corrosion.
25. The must be eye level indications of power/line in from inbuilt stabilizer.
26. Chart change and chart set options in TRCU must be available.
27. Must have MCB protection in the input power line.